

CERTIFICATE OF ANALYSIS Nr 751



Product: BioCapt® Single-Use Manufactured by Neomed srl Product Code: 79001201

Media: TRYPTONE SOYA AGAR

Lot Nr:Date of Manufacturing:Expiry date :Storage Conditions:16I220322 Sep 201622 Jun 20172-25°C up to expire date

CHEMICAL / PHYSICAL TESTS:

Appearance:	light amber coloured, clear to slightly opalescent gel forms in BioCapt device plates		
pH specification:	7.3 ± 0.2 at 25°C	pH result: 7,3	Conforms
Filling media Volume: 30 ml ± 1ml			
Media Composition: TSA- BSE free (see tecnichal data sheet cod 79001201)			

MICROBIOLOGICAL TEST CONTROLS:

Sterility Control - Absence of contamination
At the following incubation conditions (performed before gamma irradiation): $32.5 \pm 2.5 ^{\circ}\text{C aerobic for 5 days}$ Conforms

22.5 ± 2.5°C aerobic for 5 days

Fertility - Growth Promotion test (performed after gamma irradiation):

Conforms

Control strain	Medium inoculation level	Incubation Conditions	Recovery Specifications	Recovery Results	Cultural Response
E. coli ATCC 8739	10-100 viable microorganisms	24-72 h at 32.5 ± 2.5°C	70%≤R%≤200%	89,1%	Conforms
P. aeruginosa ATCC 9027	10-100 viable microorganisms	24-72 h at 32.5 ± 2.5°C	70%≤R%≤200%	80,0%	Conforms
S. aureus ATCC 6538	10-100 viable microorganisms	24-72 h at 32.5 ± 2.5°C	70%≤R%≤200%	79,2%	Conforms
B. subtilis ATCC 6633	10-100 viable microorganisms	24-72 h at 32.5 ± 2.5°C	70%≤R%≤200%	118,2%	Conforms
C. albicans ATCC 10231	10-100 viable microorganisms	72-120 h at 32.5 ± 2.5°C	70%≤R%≤200%	111,0%	Conforms
A. brasiliensis ATCC 16404	10-100 viable microorganisms	72-120 h at 32.5 ± 2.5°C	70%≤R%≤200%	72,3%	Conforms

GROWTH PROMOTION TEST IN ACCORDANCE TO HARMONISED EP – USP PHARMACOPOEIAS CHAPTER, BY QUANTITATIVE SURFACE PLATING METHOD.

RECOVERY: RESULTS AS CFU ENUMERATION ON TEST BATCH AGAR PLATES X 100/CFU ENUMERATION ON REFERENCE TESTS.

REFERENCE MEDIUM FOR RECOVERY TESTS: TRYPTIC SOYA AGAR FOR ALL BACTERIA AND SABOURAUD DEXTROSE AGAR FOR YEAST & MOLD SPECIES. **ATCC** IS REGISTERED TRADE MARK OF AMERICAN TYPE CULTURE COLLECTION.

Lot Status: Approved Lot Approval Date: 2 Nov 2016

Quality Control Manager: Dott.ssa Cinzia Bovio Signature:

NEOMED s.r.l. Via per Ossona, 22 20010 Casorezzo (MI) Italia neomed@neomed.it

tel. +39 02 93900652 +39 02 93901463 +39 02 93902434

fax. +39 02 93900968



