

CERTIFICATE OF ANALYSIS Nr 410



Product: BioCapt® Single-Use Manufactured by CPC Biotech srl Product Code: 79001201

Media: TRYPTONE SOYA AGAR

Lot Nr:Date of Manufacturing:Expiry date :Storage Conditions:18F210121 Jun 201821 Jun 20192-25°C up to expire date

CHEMICAL / PHYSICAL TESTS:

Appearance: light amber coloured, clear to slightly opalescent gel forms in BioCapt device plates

pH specification: 7.3 ± 0.2 at 25°C

pH result: 7,3

Conforms

Media Composition: TSA- BSE free (see tecnichal data sheet cod 79001201)

Conforms

MICROBIOLOGICAL TEST CONTROLS:

Sterility Control - Absence of contamination

At the following incubation conditions (performed before gamma irradiation):

 32.5 ± 2.5 °C aerobic for 7 days 22.5 ± 2.5 °C aerobic for 7 days

Conforms

Fertility - Growth Promotion test (performed after gamma irradiation):

Conforms

Control strain	Medium inoculation level	Incubation Conditions	Recovery Specifications	Recovery Results	Cultural Response
E. coli ATCC 8739	10-100 viable microorganisms	24-72 h at 32.5 ± 2.5°C	70%≤R%≤200%	84,4%	Conforms
P. aeruginosa ATCC 9027	10-100 viable microorganisms	24-72 h at 32.5 ± 2.5°C	70%≤R%≤200%	81,2%	Conforms
S. aureus ATCC 6538	10-100 viable microorganisms	24-72 h at 32.5 ± 2.5°C	70%≤R%≤200%	85,3%	Conforms
B. subtilis ATCC 6633	10-100 viable microorganisms	24-72 h at 32.5 ± 2.5°C	70%≤R%≤200%	88,9%	Conforms
C. albicans ATCC 10231	10-100 viable microorganisms	72-120 h at 32.5 ± 2.5°C	70%≤R%≤200%	91,8%	Conforms
A. brasiliensis ATCC 16404	10-100 viable microorganisms	72-120 h at 32.5 ± 2.5°C	70%≤R%≤200%	95,6%	Conforms

Growth Promotion Test in accordance to harmonised EP-USP Pharmacopoeias chapter, by quantitative surface plating method.

RECOVERY: RESULTS AS CFU ENUMERATION ON TEST BATCH AGAR PLATES X 100/CFU ENUMERATION ON REFERENCE TESTS.

REFERENCE MEDIUM FOR RECOVERY TESTS: TRYPTIC SOYA AGAR FOR ALL BACTERIA AND SABOURAUD DEXTROSE AGAR FOR YEAST & MOLD SPECIES. **ATCC** IS REGISTERED TRADE MARK OF AMERICAN TYPE CULTURE COLLECTION.

Lot status	Approved	Date 03 Aug 2018	
Quality Control	Stefania Nichele	Nichele Stefanio	
Quality Assurance	Celeste Annovazzi	Ceste Sware	

 CPC Biotech s.r.l.
 tel. +39 039 6898029
 PMS 79001201

 Via per Ossona, 22 20010 Casorezzo (MI) Italia www.cpcbiotech.it
 fax. +39 039 6899774
 Rev. 01 _18/06/2018